
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001361
9:43 24/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:30
PRESSURIZING 4:31
STERILIZING 5:52
136.0 C 221kPa 5:52
136.1 C 220kPa 6:22
136.2 C 222kPa 6:52
136.1 C 219kPa 7:22
136.4 C 223kPa 7:52
136.3 C 222kPa 8:22
136.3 C 224kPa 8:52
136.4 C 224kPa 9:22
136.4 C 224kPa 9:22
VENTING 9:22
AIR DRYING 10:19
CYCLE COMPLETE 15:17

Digital Signature #
9CAC6DB03E82B5C6
