

-----  
Midmark M11 - v1.0.5  
Total Cycles: 366

Name: RITTER2  
Sterilizer ID: V2313228  
09/15/2021 09:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.1 F 0.0  
2:00 169.2 F 2.2  
4:00 188.1 F 6.8  
6:00 208.5 F 11.8  
8:00 226.4 F 6.9  
10:00 237.9 F 12.3  
12:00 237.9 F 10.4  
14:00 237.0 F 9.4  
16:00 251.2 F 16.4  
18:00 263.1 F 23.2  
20:00 272.5 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.4 F 29.4  
2:00 271.4 F 29.0  
3:00 272.6 F 29.5  
4:00 271.6 F 28.8

Min 271.2 F 28.7  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.4 F 0.0  
5:00 229.6 F 0.0  
10:00 237.6 F 0.0  
15:00 231.2 F 0.0  
20:00 202.7 F 0.0  
25:00 233.1 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 2:02  
HEATING: 20:10  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:59:29  
CYCLE COMPLETE

09/15/2021 10:36 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01