

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000618  
13:52 30/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:36  
PRESSURIZING 2:49  
STERILIZING 3:54  
136.0 C 227kPa 3:54  
136.4 C 228kPa 4:25  
136.2 C 231kPa 4:55  
136.4 C 231kPa 5:25  
136.2 C 231kPa 5:55  
136.3 C 231kPa 6:25  
136.4 C 233kPa 6:55  
136.4 C 229kPa 7:25  
136.4 C 229kPa 7:25  
VENTING 7:25  
AIR DRYING 8:15  
CYCLE COMPLETE 23:15  
-----

Digital Signature #  
5B4F3DB96A9C4E81  
-----