
Midmark M11 - v1.0.5
Total Cycles: 1262

Name: RITTER2
Sterilizer ID: V2313228
07/06/2022 12:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.4 F 0.0
2:00 125.2 F 1.6
4:00 142.2 F 4.4
6:00 170.3 F 8.2
8:00 201.7 F 12.3
10:00 210.9 F 6.0
12:00 225.2 F 10.5
14:00 231.6 F 7.6
16:00 238.9 F 11.9
18:00 237.4 F 9.9
20:00 250.1 F 16.0
22:00 261.2 F 22.5
24:00 271.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.5 F 30.1
2:00 272.4 F 29.6
3:00 271.4 F 29.1
4:00 271.8 F 29.8

Min 271.3 F 29.1
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 206.0 F 0.0
10:00 212.9 F 0.0
15:00 218.5 F 0.0
20:00 218.6 F 0.0
25:00 216.0 F 0.0
30:00 213.2 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:40
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:02:49
CYCLE COMPLETE

07/06/2022 01:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____