
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000016
12:00 08/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:45
PRESSURIZING 1:50
STERILIZING 2:36
136.3 C 223kPa 2:36
136.0 C 217kPa 3:06
136.4 C 223kPa 3:36
136.4 C 226kPa 4:06
136.4 C 225kPa 4:36
136.4 C 225kPa 5:06
135.9 C 219kPa 5:36
136.1 C 219kPa 6:06
136.1 C 219kPa 6:06
VENTING 6:06
AIR DRYING 6:54
CYCLE COMPLETE 16:28

Digital Signature #
187D07A8285B02F5
