
Midmark M11 - v1.0.5
Total Cycles: 2997

Name: RITTER2
Sterilizer ID: V2313228
10/11/2023 12:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.5 F 0.0
2:00 186.2 F 3.7
4:00 201.5 F 8.8
6:00 214.2 F 8.2
8:00 226.5 F 9.8
10:00 232.0 F 8.3
12:00 242.8 F 12.7
14:00 244.1 F 13.0
16:00 257.3 F 20.0
18:00 269.0 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.7 F 29.9
2:00 271.8 F 29.1
3:00 272.4 F 30.1
4:00 272.4 F 29.6

Min 271.6 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.5 F 0.0
5:00 224.0 F 0.0
10:00 221.0 F 0.0
15:00 217.1 F 0.0
20:00 213.9 F 0.0
25:00 216.8 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 19:06
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:01
CYCLE COMPLETE

10/11/2023 01:46 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____