
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001572
11:44 08/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:52
STERILIZING 3:54
136.4 C 222kPa 3:54
136.2 C 223kPa 4:24
136.4 C 225kPa 4:54
136.3 C 223kPa 5:24
136.4 C 222kPa 5:54
136.2 C 222kPa 6:24
136.4 C 224kPa 6:54
136.4 C 224kPa 7:24
136.4 C 224kPa 7:24
VENTING 7:24
AIR DRYING 8:15
CYCLE COMPLETE 8:20

Digital Signature #
08B29E669FBFA381
