
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001271
14:38 26/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:24
STERILIZING 4:58
136.4 C 224kPa 4:58
136.4 C 224kPa 5:28
136.3 C 224kPa 5:58
136.3 C 223kPa 6:28
136.4 C 225kPa 6:58
136.2 C 223kPa 7:28
136.1 C 223kPa 7:58
136.4 C 225kPa 8:28
136.4 C 225kPa 8:28
VENTING 8:28
AIR DRYING 9:24
CYCLE COMPLETE 12:40

Digital Signature #
A6E26A5DB9FC2D7B
