
Midmark M11 - v1.0.0
Total Cycles: 1630

Name: RITTER 1
Sterilizer ID: V2178198
02/14/2022 02:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.8 F 0.0
2:00 137.4 F 1.4
4:00 156.7 F 4.3
6:00 189.9 F 8.5
8:00 205.2 F 11.2
10:00 220.7 F 6.8
12:00 233.4 F 11.9
14:00 234.9 F 8.9
16:00 240.9 F 9.9
18:00 247.6 F 13.9
20:00 260.3 F 20.9
22:00 271.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.0 F 28.5
2:00 272.5 F 29.0
3:00 271.5 F 28.1
4:00 272.6 F 28.8

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 218.6 F 0.0
10:00 231.0 F 0.0
15:00 228.8 F 0.0
20:00 225.7 F 0.0
25:00 226.9 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 22:32
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:00:43
CYCLE COMPLETE

02/14/2022 03:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____