
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
2.4uS / 1.5ppm
CYCLE NUMBER 000014
14:22 13/01/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:11
PRESSURIZING 6:13
STERILIZING 8:03
136.1 C 224kPa 8:03
136.3 C 226kPa 8:33
136.5 C 228kPa 9:03
136.4 C 227kPa 9:33
136.4 C 228kPa 10:03
136.4 C 227kPa 10:33
136.4 C 227kPa 11:04
136.4 C 228kPa 11:33
VENTING 11:33
AIR DRYING 12:50
CYCLE COMPLETE 42:50

Digital Signature #
A3B0858C45BC8DA5
