
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006953
8:01 16/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:33
STERILIZING 3:31
136.0 C 220kPa 3:31
136.4 C 220kPa 4:01
136.3 C 219kPa 4:31
136.4 C 222kPa 5:01
136.4 C 221kPa 5:31
136.3 C 221kPa 6:01
136.4 C 223kPa 6:31
136.3 C 220kPa 7:01
136.3 C 220kPa 7:01
VENTING 7:01
AIR DRYING 7:49
CYCLE COMPLETE 37:50

Digital Signature #
DABC8FFC23898406
