
Midmark M11 - v1.0.0
Total Cycles: 3887

Name: RITTER 1
Sterilizer ID: V2178198
10/11/2023 09:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.2 F 0.0
2:00 174.9 F 2.1
4:00 181.2 F 5.8
6:00 209.9 F 11.0
8:00 220.3 F 5.7
10:00 233.3 F 11.0
12:00 232.4 F 8.0
14:00 241.4 F 11.1
16:00 245.5 F 13.2
18:00 258.2 F 19.9
20:00 269.5 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.3 F 28.9
2:00 271.6 F 28.8
3:00 272.2 F 28.8
4:00 271.5 F 28.7

Min 271.3 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 211.5 F 0.0
10:00 234.0 F 0.0
15:00 227.1 F 0.0
20:00 228.1 F 0.0
25:00 229.2 F 0.0
30:00 229.9 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 20:59
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:00:04
CYCLE COMPLETE

10/11/2023 10:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____