

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000216  
10:21 22/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:53  
PRESSURIZING 3:55  
STERILIZING 5:20  
136.4 C 233kPa 5:20  
136.4 C 233kPa 5:50  
136.4 C 233kPa 6:20  
136.3 C 233kPa 6:50  
136.4 C 233kPa 7:20  
136.2 C 231kPa 7:50  
136.3 C 233kPa 8:20  
136.4 C 233kPa 8:50  
136.4 C 233kPa 8:50  
VENTING 8:50  
AIR DRYING 9:52  
CYCLE COMPLETE 21:15  
-----

Digital Signature #  
D5A0CF5F627ACEEA  
-----