
Midmark M11 - v1.0.0
Total Cycles: 585

Name: MIDMARK
Sterilizer ID: V2178198
12/29/2020 11:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.6 F 0.0
2:00 127.3 F 0.8
4:00 149.5 F 3.1
6:00 179.6 F 6.9
8:00 205.2 F 12.6
10:00 217.6 F 5.9
12:00 230.0 F 10.9
14:00 231.5 F 7.1
16:00 244.3 F 13.1
18:00 242.7 F 11.5
20:00 255.6 F 18.1
22:00 267.1 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.3 F 28.5
2:00 271.9 F 28.5
3:00 272.0 F 28.3
4:00 272.1 F 28.7

Min 271.4 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.1 F 0.0
5:00 206.5 F 0.0
10:00 199.4 F 0.0
15:00 197.5 F 0.0
20:00 195.1 F 0.0
25:00 194.0 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:25
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:01:30
CYCLE COMPLETE

12/29/2020 12:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____