
Midmark M11 - v1.0.5
Total Cycles: 1760

Name: RITTER2
Sterilizer ID: V2313228
11/22/2022 12:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.7 F 0.0
2:00 200.4 F 3.4
4:00 207.2 F 9.4
6:00 223.3 F 7.7
8:00 235.3 F 12.0
10:00 237.3 F 10.6
12:00 237.6 F 9.8
14:00 252.6 F 17.3
16:00 265.2 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.5 F 29.7
2:00 271.5 F 29.0
3:00 272.5 F 29.8
4:00 272.0 F 29.3

Min 271.4 F 29.0
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 208.5 F 0.0
10:00 220.2 F 0.0
15:00 223.7 F 0.0
20:00 223.2 F 0.0
25:00 223.2 F 0.0
30:00 222.6 F 0.0

Drying Complete

FILLING: 2:48
HEATING: 17:43
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:44
CYCLE COMPLETE

11/22/2022 01:07 PM
ACCEPTED: X
REJECTED:
Operator: U02