
Midmark M11 - v1.0.0
Total Cycles: 3130

Name: RITTER 1
Sterilizer ID: V2178198
04/19/2023 12:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.5 F 0.0
2:00 200.3 F 1.3
4:00 220.2 F 7.0
6:00 229.7 F 12.5
8:00 233.4 F 9.1
10:00 239.1 F 9.5
12:00 244.7 F 12.4
14:00 246.2 F 13.3
16:00 259.3 F 20.4
18:00 270.9 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.6 F 28.8
2:00 271.6 F 28.2
3:00 272.8 F 29.0
4:00 272.1 F 28.4

Min 271.6 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 225.0 F 0.0
10:00 234.3 F 0.0
15:00 229.4 F 0.0
20:00 226.3 F 0.0
25:00 225.3 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 2:14
HEATING: 18:44
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

04/19/2023 01:33 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____