
Midmark M11 - v1.0.0
Total Cycles: 4259

Name: RITTER 1
Sterilizer ID: V2178198
02/01/2024 01:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.7 F 0.0
2:00 122.8 F 1.6
4:00 139.2 F 4.2
6:00 178.5 F 7.8
8:00 195.7 F 12.8
10:00 216.9 F 5.2
12:00 222.8 F 9.7
14:00 234.3 F 9.1
16:00 239.9 F 11.0
18:00 236.0 F 8.8
20:00 249.9 F 15.3
22:00 261.5 F 21.7
24:00 272.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 271.8 F 28.8
2:00 271.8 F 28.7
3:00 272.2 F 28.8
4:00 272.1 F 28.9

Min 271.6 F 28.7
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.6 F 0.0
5:00 243.0 F 0.0
10:00 239.3 F 0.0
15:00 233.9 F 0.0
20:00 230.9 F 0.0
25:00 206.8 F 0.0
30:00 195.5 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 25:10
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:03:38
CYCLE COMPLETE

02/01/2024 02:40 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____