
Midmark M11 - v1.0.0
Total Cycles: 3796

Name: RITTER 1
Sterilizer ID: V2178198
09/19/2023 02:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.2 F 0.0
2:00 179.1 F 2.4
4:00 195.6 F 7.0
6:00 215.8 F 11.9
8:00 225.7 F 6.6
10:00 233.8 F 11.3
12:00 234.0 F 8.0
14:00 244.3 F 12.2
16:00 245.0 F 12.5
18:00 257.0 F 18.9
20:00 267.8 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.3 F 28.6
2:00 271.5 F 28.4
3:00 272.3 F 28.6
4:00 271.3 F 28.3

Min 271.2 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 227.4 F 0.0
10:00 221.0 F 0.0
15:00 221.6 F 0.0
20:00 222.2 F 0.0
25:00 226.2 F 0.0
30:00 220.2 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 21:21
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:00:35
CYCLE COMPLETE

09/19/2023 03:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____