

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002090  
9:14 24/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:22  
PRESSURIZING 3:27  
STERILIZING 4:32  
136.3 C 223kPa 4:32  
136.4 C 223kPa 5:03  
136.4 C 224kPa 5:33  
136.3 C 225kPa 6:03  
136.4 C 224kPa 6:33  
136.4 C 225kPa 7:03  
136.3 C 222kPa 7:33  
136.4 C 225kPa 8:03  
136.4 C 225kPa 8:03  
VENTING 8:03  
AIR DRYING 8:56  
CYCLE COMPLETE 9:09  
-----

Digital Signature #  
080C766E00AC9D23  
-----