
Midmark M11 - v1.0.0
Total Cycles: 4338

Name: RITTER 1
Sterilizer ID: V2178198
02/27/2024 02:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 87.1 F 0.0
2:00 87.5 F 0.5
4:00 104.2 F 2.2
6:00 139.5 F 4.5
8:00 169.1 F 6.7
10:00 189.8 F 10.5
12:00 210.3 F 6.8
14:00 210.7 F 4.9
16:00 220.9 F 8.3
18:00 231.0 F 12.1
20:00 231.8 F 7.9
22:00 242.1 F 12.8
24:00 239.9 F 10.7
26:00 251.4 F 16.3
28:00 261.7 F 22.2
30:00 271.4 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.6
1:00 272.2 F 29.2
2:00 271.1 F 28.5
3:00 272.2 F 29.3
4:00 271.4 F 28.6

Min 271.0 F 28.4
Max 272.3 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 210.7 F 0.0
10:00 203.1 F 0.0
15:00 202.3 F 0.0
20:00 216.4 F 0.0
25:00 224.6 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 30:40
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:09:21
CYCLE COMPLETE

02/27/2024 03:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____