
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001655
16:48 06/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:11
STERILIZING 3:11
136.4 C 224kPa 3:11
136.3 C 224kPa 3:41
136.3 C 224kPa 4:11
136.4 C 222kPa 4:41
136.4 C 226kPa 5:11
136.1 C 222kPa 5:41
136.4 C 224kPa 6:11
136.2 C 222kPa 6:41
136.2 C 222kPa 6:41
VENTING 6:41
AIR DRYING 7:33
CYCLE COMPLETE 7:36

Digital Signature #
91915776AB2B167E
