
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001563
14:35 06/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:21
STERILIZING 4:25
136.5 C 225kPa 4:25
136.4 C 226kPa 4:55
136.2 C 223kPa 5:25
136.0 C 222kPa 5:55
136.4 C 226kPa 6:25
136.3 C 224kPa 6:55
136.2 C 224kPa 7:25
136.2 C 224kPa 7:55
136.2 C 224kPa 7:55
VENTING 7:56
AIR DRYING 8:51
CYCLE COMPLETE 11:31

Digital Signature #
492F45BB669E3D6E
