
Midmark M11 - v1.0.0
Total Cycles: 2993

Name: RITTER 1
Sterilizer ID: V2178198
03/13/2023 07:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.4 F 0.0
2:00 77.5 F 0.0
4:00 96.3 F 0.0
6:00 118.9 F 0.0
8:00 150.6 F 1.2
10:00 186.7 F 3.2
12:00 201.9 F 5.9
14:00 215.8 F 9.2
16:00 227.5 F 12.8
18:00 228.6 F 7.1
20:00 236.7 F 11.5
22:00 234.2 F 8.1
24:00 244.8 F 12.4
26:00 243.9 F 12.1
28:00 255.7 F 18.2
30:00 266.1 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.4 F 28.8
2:00 271.3 F 28.2
3:00 272.4 F 28.8
4:00 271.3 F 28.0

Min 271.3 F 28.0
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 231.7 F 0.0
10:00 224.2 F 0.0
15:00 231.0 F 0.0
20:00 230.6 F 0.0
25:00 227.5 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:41
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:09:41
CYCLE COMPLETE

03/13/2023 08:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____