

-----  
Midmark M11 - v1.0.0  
Total Cycles: 723

Name: RITTER 1  
Sterilizer ID: V2178198  
03/17/2021 09:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.4 F 0.0  
2:00 185.6 F 3.0  
4:00 202.7 F 8.2  
6:00 230.5 F 8.0  
8:00 229.8 F 11.6  
10:00 238.8 F 9.6  
12:00 235.0 F 7.9  
14:00 251.9 F 15.9  
16:00 264.8 F 23.4  
18:00 273.0 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.5  
1:00 272.4 F 28.5  
2:00 272.5 F 28.7  
3:00 272.0 F 28.8  
4:00 272.5 F 28.6  
  
Min 271.8 F 28.4  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.4 F 0.0  
5:00 241.7 F 0.0  
10:00 238.9 F 0.0  
15:00 175.8 F 0.0  
20:00 169.2 F 0.0  
25:00 164.7 F 0.0  
30:00 165.7 F 0.0  
Drying Complete

FILLING: 1:48  
HEATING: 18:11  
STERILIZING: 4:00  
VENTING: 2:56  
DRYING: 30:00  
TOTAL CYCLE: 00:56:55  
CYCLE COMPLETE

03/17/2021 09:59 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01