
Midmark M11 - v1.0.0
Total Cycles: 5045

Name: RITTER 1
Sterilizer ID: V2178198
11/01/2024 11:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.1 F 0.0
2:00 172.6 F 0.5
4:00 188.1 F 3.7
6:00 207.1 F 8.4
8:00 222.7 F 13.2
10:00 227.5 F 7.9
12:00 239.6 F 13.0
14:00 239.1 F 10.4
16:00 235.8 F 8.3
18:00 250.2 F 15.2
20:00 261.7 F 21.8
22:00 272.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.2
2:00 272.0 F 28.6
3:00 271.4 F 28.6
4:00 272.4 F 29.2

Min 271.2 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.0 F 0.0
5:00 250.2 F 0.0
10:00 196.6 F 0.0
15:00 192.1 F 0.0
20:00 192.0 F 0.0
25:00 226.6 F 0.0
30:00 196.7 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 22:35
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:01:13
CYCLE COMPLETE

11/01/2024 12:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____