
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000038
10:01 15/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:26
PRESSURIZING 3:27
STERILIZING 4:29
136.5 C 224kPa 4:29
136.4 C 225kPa 4:59
136.2 C 224kPa 5:29
136.3 C 224kPa 5:59
136.3 C 222kPa 6:29
136.0 C 223kPa 6:59
136.4 C 225kPa 7:29
136.4 C 224kPa 7:59
136.4 C 224kPa 7:59
VENTING 7:59
AIR DRYING 8:53
CYCLE COMPLETE 9:00

Digital Signature #
E9EEF3990D4F68BE
