

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2655

Name: RITTER 1  
Sterilizer ID: V2178198  
12/07/2022 04:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.2 F 0.0  
2:00 173.8 F 3.7  
4:00 189.0 F 8.0  
6:00 212.4 F 13.1  
8:00 222.3 F 7.2  
10:00 235.0 F 12.6  
12:00 236.6 F 9.7  
14:00 236.4 F 8.5  
16:00 249.6 F 15.1  
18:00 262.2 F 22.2  
20:00 272.4 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.4 F 28.9  
2:00 271.7 F 28.6  
3:00 272.4 F 28.9  
4:00 271.5 F 28.4

Min 271.5 F 28.3  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.4 F 0.0  
5:00 222.5 F 0.0  
10:00 235.8 F 0.0  
15:00 228.0 F 0.0  
20:00 230.6 F 0.0  
25:00 226.6 F 0.0  
30:00 224.0 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 20:14  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:30  
CYCLE COMPLETE

12/07/2022 05:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_