
Midmark M11 - v1.0.0
Total Cycles: 3498

Name: RITTER 1
Sterilizer ID: V2178198
07/06/2023 02:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.4 F 0.0
2:00 161.4 F 0.0
4:00 186.1 F 1.7
6:00 200.9 F 6.2
8:00 216.3 F 11.1
10:00 224.9 F 5.8
12:00 234.8 F 11.2
14:00 234.5 F 8.3
16:00 241.1 F 10.8
18:00 246.6 F 13.4
20:00 259.0 F 20.1
22:00 270.2 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 272.1 F 28.8
2:00 271.8 F 28.4
3:00 271.5 F 28.4
4:00 272.5 F 28.8

Min 271.2 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 204.7 F 0.0
10:00 225.4 F 0.0
15:00 234.2 F 0.0
20:00 239.7 F 0.0
25:00 227.7 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 22:45
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:55
CYCLE COMPLETE

07/06/2023 03:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____