
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 005071
13:38 06/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:35
STERILIZING 3:33
136.4 C 223kPa 3:33
136.3 C 223kPa 4:03
136.4 C 223kPa 4:33
136.3 C 225kPa 5:03
136.4 C 224kPa 5:33
136.3 C 223kPa 6:03
136.2 C 223kPa 6:33
136.3 C 222kPa 7:03
136.3 C 222kPa 7:03
VENTING 7:03
AIR DRYING 7:58
CYCLE COMPLETE 10:20

Digital Signature #
69199E7754B60883
