

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1685

Name: RITTER2  
Sterilizer ID: V2313228  
11/07/2022 08:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.3 F 0.0  
2:00 82.0 F 0.0  
4:00 127.1 F 0.0  
6:00 163.8 F 0.1  
8:00 178.4 F 0.3  
10:00 178.2 F 0.5  
12:00 197.3 F 1.0  
14:00 216.1 F 1.9  
16:00 227.7 F 5.5  
18:00 236.8 F 9.2  
20:00 246.0 F 13.5  
22:00 242.3 F 11.5  
24:00 239.7 F 10.3  
26:00 237.7 F 9.5  
28:00 252.3 F 16.5  
30:00 264.1 F 23.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.5  
1:00 272.4 F 28.7  
2:00 272.7 F 29.3  
3:00 271.7 F 28.5  
4:00 272.5 F 28.8

Min 271.7 F 28.3  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.3 F 0.0  
5:00 218.5 F 0.0  
10:00 225.3 F 0.0  
15:00 219.2 F 0.0  
20:00 220.3 F 0.0  
25:00 220.3 F 0.0  
30:00 219.3 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 31:57  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:09:53  
CYCLE COMPLETE

11/07/2022 09:43 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_