

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3981

Name: RITTER 1  
Sterilizer ID: V2178198  
11/06/2023 09:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.7 F 0.0  
2:00 69.9 F 0.2  
4:00 83.4 F 1.2  
6:00 120.8 F 3.4  
8:00 143.8 F 6.2  
10:00 166.6 F 9.7  
12:00 188.1 F 13.4  
14:00 202.9 F 3.6  
16:00 214.8 F 6.7  
18:00 225.3 F 10.1  
20:00 233.2 F 9.7  
22:00 235.8 F 9.6  
24:00 239.6 F 10.0  
26:00 244.4 F 12.5  
28:00 256.0 F 18.4  
30:00 266.3 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.6 F 29.0  
2:00 271.6 F 28.1  
3:00 272.4 F 29.0  
4:00 272.2 F 28.6

Min 271.3 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.7 F 0.0  
5:00 191.3 F 0.0  
10:00 205.2 F 0.0  
15:00 196.0 F 0.0  
20:00 207.9 F 0.0  
25:00 221.2 F 0.0  
30:00 222.5 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 31:45  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:09:58  
CYCLE COMPLETE

11/06/2023 10:14 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_