

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3669

Name: RITTER2  
Sterilizer ID: V2313228  
06/05/2024 10:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 103.8 F 0.0  
2:00 103.7 F 0.5  
4:00 119.9 F 2.2  
6:00 146.9 F 5.1  
8:00 172.1 F 8.7  
10:00 195.2 F 13.1  
12:00 219.3 F 6.0  
14:00 232.5 F 10.1  
16:00 232.1 F 8.9  
18:00 238.9 F 10.9  
20:00 234.1 F 8.3  
22:00 247.9 F 14.5  
24:00 258.7 F 20.6  
26:00 269.0 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 271.9 F 29.1  
2:00 272.1 F 29.5  
3:00 272.1 F 29.6  
4:00 271.5 F 28.9

Min 271.3 F 28.9  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.0 F 0.0  
5:00 233.9 F 0.0  
10:00 224.2 F 0.0  
15:00 218.6 F 0.0  
20:00 223.0 F 0.0  
25:00 230.8 F 0.0  
30:00 228.4 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 27:05  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:05:14  
CYCLE COMPLETE

06/05/2024 11:59 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_