

-----  
Midmark M11 - v1.0.5  
Total Cycles: 481

Name: RITTER2  
Sterilizer ID: V2313228  
10/25/2021 02:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.2 F 0.0  
2:00 189.1 F 3.2  
4:00 191.4 F 8.9  
6:00 213.5 F 7.3  
8:00 224.9 F 11.4  
10:00 234.1 F 9.8  
12:00 236.0 F 8.7  
14:00 250.8 F 16.2  
16:00 262.8 F 23.5  
18:00 271.3 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 273.4 F 30.2  
1:00 272.5 F 30.4  
2:00 272.4 F 30.5  
3:00 272.5 F 30.3  
4:00 272.5 F 30.2  
  
Min 271.9 F 30.0  
Max 273.4 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.4 F 0.0  
5:00 235.6 F 0.0  
10:00 224.8 F 0.0  
15:00 220.6 F 0.0  
20:00 215.8 F 0.0  
25:00 211.9 F 0.0  
30:00 214.2 F 0.0  
Drying Complete

FILLING: 2:06  
HEATING: 18:40  
STERILIZING: 4:00  
VENTING: 2:60  
DRYING: 30:00  
TOTAL CYCLE: 00:57:46  
CYCLE COMPLETE

10/25/2021 02:59 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01