
Midmark M11 - v1.0.5
Total Cycles: 4382

Name: RITTER2
Sterilizer ID: V2313228
02/26/2025 12:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.0 F 0.0
2:00 159.6 F 2.6
4:00 172.2 F 6.8
6:00 196.4 F 11.6
8:00 215.6 F 5.5
10:00 224.5 F 10.0
12:00 234.1 F 9.0
14:00 240.9 F 12.5
16:00 240.5 F 11.3
18:00 252.7 F 17.5
20:00 263.5 F 23.9
22:00 272.8 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.6
1:00 272.6 F 30.1
2:00 271.5 F 29.6
3:00 272.4 F 30.1
4:00 271.6 F 29.5

Min 271.4 F 29.5
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 196.9 F 0.0
10:00 187.0 F 0.0
15:00 176.4 F 0.0
20:00 172.8 F 0.0
25:00 174.8 F 0.0
30:00 172.5 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 22:17
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:02:21
CYCLE COMPLETE

02/26/2025 01:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____