
Midmark M11 - v1.0.5
Total Cycles: 689

Name: RITTER2
Sterilizer ID: V2313228
01/12/2022 02:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.5 F 0.0
2:00 192.0 F 3.8
4:00 204.4 F 9.3
6:00 222.2 F 7.3
8:00 233.8 F 11.6
10:00 235.7 F 9.9
12:00 236.0 F 8.9
14:00 250.5 F 16.2
16:00 262.7 F 23.4
18:00 272.1 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.4 F 29.6
2:00 272.2 F 30.1
3:00 271.8 F 29.2
4:00 272.5 F 30.0

Min 271.4 F 29.2
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 222.3 F 0.0
10:00 229.6 F 0.0
15:00 227.8 F 0.0
20:00 227.1 F 0.0
25:00 227.2 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 18:07
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:25
CYCLE COMPLETE

01/12/2022 03:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____