
Midmark M11 - v1.0.5
Total Cycles: 4477

Name: RITTER2
Sterilizer ID: V2313228
04/02/2025 10:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.1 F 0.0
2:00 156.3 F 1.6
4:00 157.8 F 4.5
6:00 182.7 F 8.3
8:00 195.8 F 12.8
10:00 225.3 F 6.8
12:00 232.2 F 11.4
14:00 232.2 F 7.8
16:00 245.0 F 13.3
18:00 242.9 F 11.8
20:00 254.5 F 17.9
22:00 264.8 F 24.2
24:00 270.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.8
1:00 271.8 F 29.9
2:00 271.7 F 30.0
3:00 271.6 F 29.9
4:00 271.6 F 29.8

Min 271.5 F 29.7
Max 272.4 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 183.0 F 0.0
10:00 190.0 F 0.0
15:00 194.8 F 0.0
20:00 193.1 F 0.0
25:00 191.7 F 0.0
30:00 187.6 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 24:44
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:03:38
CYCLE COMPLETE

04/02/2025 11:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____