

-----  
Midmark M11 - v1.0.0  
Total Cycles: 628

Name: MIDMARK  
Sterilizer ID: V2178198  
01/26/2021 08:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 140.5 F 0.0  
2:00 141.2 F 1.1  
4:00 155.6 F 3.7  
6:00 181.0 F 7.6  
8:00 202.4 F 12.6  
10:00 216.2 F 5.5  
12:00 231.2 F 10.5  
14:00 233.0 F 7.5  
16:00 244.6 F 13.3  
18:00 243.1 F 11.8  
20:00 255.8 F 18.3  
22:00 267.0 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.3 F 28.6  
2:00 272.3 F 28.8  
3:00 272.1 F 28.4  
4:00 272.4 F 28.9

Min 271.6 F 28.1  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.0 F 0.0  
5:00 233.2 F 0.0  
10:00 230.6 F 0.0  
15:00 229.9 F 0.0  
20:00 214.7 F 0.0  
25:00 208.0 F 0.0  
30:00 206.3 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 23:25  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:01:30  
CYCLE COMPLETE

01/26/2021 09:43 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01