
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001132
9:47 05/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:08
STERILIZING 3:14
136.5 C 220kPa 3:14
136.2 C 222kPa 3:44
CYCLE ABORTED 4:03
STOP BUTTON PRESSED
NOT STERILE
END TIME 4:03

Digital Signature #
A65FAE020BE90A63
