

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002495  
9:10 10/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:00  
PRESSURIZING 2:04  
STERILIZING 3:13  
136.0 C 220kPa 3:13  
136.4 C 225kPa 3:43  
136.4 C 225kPa 4:13  
136.3 C 223kPa 4:43  
136.4 C 224kPa 5:13  
136.4 C 223kPa 5:43  
136.3 C 223kPa 6:13  
136.4 C 225kPa 6:43  
136.4 C 225kPa 6:43  
VENTING 6:43  
AIR DRYING 7:35  
CYCLE COMPLETE 28:42  
-----

Digital Signature #  
5C870F6CC56DFDCE  
-----