

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2458

Name: RITTER 1  
Sterilizer ID: V2178198  
10/19/2022 12:28 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.8 F 0.0  
2:00 173.1 F 4.0  
4:00 187.8 F 8.2  
6:00 207.5 F 13.0  
8:00 217.6 F 6.9  
10:00 231.5 F 12.0  
12:00 233.3 F 8.7  
14:00 239.8 F 10.3  
16:00 245.4 F 13.3  
18:00 257.7 F 19.8  
20:00 269.0 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.4 F 29.3  
2:00 271.7 F 28.6  
3:00 271.9 F 29.2  
4:00 272.1 F 29.1  
  
Min 271.3 F 28.5  
Max 272.4 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.2 F 0.0  
5:00 214.1 F 0.0  
10:00 231.0 F 0.0  
15:00 231.3 F 0.0  
20:00 227.0 F 0.0  
25:00 222.7 F 0.0  
30:00 222.6 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 21:07  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:59:35  
CYCLE COMPLETE

10/19/2022 01:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_