
Midmark M11 - v1.0.5
Total Cycles: 2877

Name: RITTER2
Sterilizer ID: V2313228
09/13/2023 12:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.2 F 0.0
2:00 153.9 F 0.2
4:00 169.0 F 3.1
6:00 190.0 F 6.7
8:00 204.6 F 10.8
10:00 217.5 F 5.8
12:00 232.5 F 9.1
14:00 236.2 F 11.6
16:00 240.8 F 11.3
18:00 239.0 F 10.2
20:00 252.1 F 16.6
22:00 263.1 F 23.1
24:00 272.2 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.9 F 29.5
2:00 271.8 F 28.8
3:00 272.2 F 29.5
4:00 272.6 F 29.3

Min 271.4 F 28.7
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.2 F 0.0
5:00 233.0 F 0.0
10:00 220.9 F 0.0
15:00 217.7 F 0.0
20:00 208.7 F 0.0
25:00 212.2 F 0.0
30:00 215.2 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:10
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:02:42
CYCLE COMPLETE

09/13/2023 01:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____