
Midmark M11 - v1.0.5
Total Cycles: 4274

Name: RITTER2
Sterilizer ID: V2313228
01/17/2025 11:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.2 F 0.0
2:00 154.2 F 0.1
4:00 159.9 F 0.3
6:00 189.9 F 1.0
8:00 207.0 F 4.4
10:00 220.1 F 8.4
12:00 232.5 F 12.8
14:00 232.6 F 9.2
16:00 240.5 F 10.9
18:00 243.8 F 13.0
20:00 243.4 F 12.2
22:00 255.3 F 18.4
24:00 265.8 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.7
1:00 271.8 F 29.1
2:00 272.5 F 29.2
3:00 271.7 F 28.6
4:00 271.8 F 29.2

Min 271.3 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.6 F 0.0
5:00 183.9 F 0.0
10:00 190.0 F 0.0
15:00 189.9 F 0.0
20:00 187.6 F 0.0
25:00 185.4 F 0.0
30:00 183.5 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:43
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:04:36
CYCLE COMPLETE

01/17/2025 12:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____