
Midmark M11 - v1.0.5
Total Cycles: 2434

Name: RITTER2
Sterilizer ID: V2313228
05/30/2023 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.0 F 0.0
2:00 172.8 F 3.0
4:00 173.7 F 7.0
6:00 198.7 F 12.0
8:00 216.2 F 6.5
10:00 230.8 F 11.9
12:00 233.4 F 8.9
14:00 239.8 F 10.1
16:00 245.9 F 13.7
18:00 258.2 F 20.4
20:00 269.0 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.5 F 29.8
2:00 271.5 F 29.0
3:00 272.2 F 30.0
4:00 272.3 F 29.5

Min 271.3 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.8 F 0.0
5:00 216.3 F 0.0
10:00 231.3 F 0.0
15:00 237.7 F 0.0
20:00 223.9 F 0.0
25:00 219.1 F 0.0
30:00 215.5 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 21:07
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:30
CYCLE COMPLETE

05/30/2023 11:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____