
Midmark M11 - v1.0.0
Total Cycles: 3626

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2023 12:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.6 F 0.0
2:00 181.3 F 3.7
4:00 192.3 F 8.0
6:00 214.0 F 13.4
8:00 222.4 F 8.0
10:00 235.3 F 13.2
12:00 237.3 F 10.5
14:00 234.8 F 8.4
16:00 249.2 F 15.1
18:00 260.6 F 21.5
20:00 271.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.4 F 29.1
2:00 271.3 F 28.5
3:00 272.5 F 29.2
4:00 271.6 F 28.5

Min 271.3 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 216.4 F 0.0
10:00 209.4 F 0.0
15:00 200.7 F 0.0
20:00 194.3 F 0.0
25:00 189.8 F 0.0
30:00 185.5 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 20:42
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:59:45
CYCLE COMPLETE

08/09/2023 01:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____