

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001610  
10:18 11/24/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:38  
PRESSURIZING 6:28  
STERILIZING 7:51  
136.0 C 220kPa 7:51  
136.1 C 219kPa 8:21  
136.4 C 220kPa 8:51  
136.4 C 222kPa 9:21  
136.4 C 222kPa 9:51  
136.3 C 218kPa 10:21  
136.2 C 221kPa 10:51  
136.3 C 222kPa 11:21  
VENTING 11:21  
AIR DRYING 12:24  
CYCLE COMPLETE 29:09

-----  
Digital Signature #  
09FEAA3FC30C27A9  
-----