

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002538  
9:04 26/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:53  
PRESSURIZING 2:02  
STERILIZING 3:12  
136.5 C 224kPa 3:12  
136.4 C 224kPa 3:42  
136.4 C 224kPa 4:12  
136.4 C 225kPa 4:42  
136.3 C 224kPa 5:12  
136.3 C 224kPa 5:42  
136.4 C 225kPa 6:12  
136.2 C 223kPa 6:42  
136.2 C 223kPa 6:42  
VENTING 6:42  
AIR DRYING 7:32  
CYCLE COMPLETE 35:31  
-----

Digital Signature #  
B16600B971629CAD  
-----