

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2466

Name: RITTER 1  
Sterilizer ID: V2178198  
10/20/2022 01:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.8 F 0.0  
2:00 162.3 F 3.6  
4:00 181.8 F 8.0  
6:00 202.1 F 13.0  
8:00 224.4 F 7.8  
10:00 239.2 F 13.1  
12:00 238.8 F 10.3  
14:00 234.7 F 8.1  
16:00 250.2 F 15.2  
18:00 262.2 F 22.1  
20:00 272.8 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 271.5 F 28.7  
2:00 271.8 F 28.7  
3:00 272.4 F 28.9  
4:00 272.8 F 29.2  
  
Min 271.2 F 28.7  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.5 F 0.0  
5:00 231.0 F 0.0  
10:00 228.9 F 0.0  
15:00 233.2 F 0.0  
20:00 224.6 F 0.0  
25:00 227.2 F 0.0  
30:00 229.3 F 0.0  
Drying Complete

FILLING: 1:59  
HEATING: 20:58  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:00:13  
CYCLE COMPLETE

10/20/2022 02:26 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_