
Midmark M11 - v1.0.0
Total Cycles: 2099

Name: RITTER 1
Sterilizer ID: V2178198
07/07/2022 11:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.8 F 0.0
2:00 171.5 F 2.7
4:00 188.4 F 7.3
6:00 209.8 F 12.6
8:00 221.6 F 6.3
10:00 232.6 F 11.3
12:00 231.7 F 7.9
14:00 242.9 F 11.7
16:00 243.9 F 12.5
18:00 256.7 F 19.2
20:00 267.9 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.7
1:00 272.4 F 29.3
2:00 271.5 F 28.6
3:00 272.5 F 29.2
4:00 271.5 F 28.6

Min 271.3 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 205.2 F 0.0
10:00 216.6 F 0.0
15:00 219.1 F 0.0
20:00 214.9 F 0.0
25:00 220.4 F 0.0
30:00 217.4 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 21:10
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

07/07/2022 12:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____