
Midmark M11 - v1.0.5
Total Cycles: 1564

Name: RITTER2
Sterilizer ID: V2313228
10/04/2022 03:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.4 F 0.0
2:00 193.1 F 0.6
4:00 208.8 F 5.6
6:00 225.7 F 11.1
8:00 230.4 F 8.1
10:00 239.9 F 11.4
12:00 244.5 F 13.2
14:00 245.0 F 12.9
16:00 257.9 F 20.0
18:00 269.8 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.1 F 29.4
2:00 272.5 F 29.4
3:00 272.5 F 29.4
4:00 272.3 F 29.6

Min 270.9 F 28.4
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 224.8 F 0.0
10:00 242.2 F 0.0
15:00 233.5 F 0.0
20:00 223.0 F 0.0
25:00 221.4 F 0.0
30:00 220.5 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 19:03
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:01
CYCLE COMPLETE

10/04/2022 04:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____