
Midmark M11 - v1.0.5
Total Cycles: 970

Name: RITTER2
Sterilizer ID: V2313228
04/07/2022 01:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.6 F 0.0
2:00 150.7 F 0.3
4:00 181.9 F 3.3
6:00 202.6 F 7.9
8:00 229.5 F 13.2
10:00 232.3 F 9.3
12:00 235.8 F 9.6
14:00 244.7 F 13.5
16:00 244.7 F 12.9
18:00 257.8 F 20.0
20:00 269.6 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.5 F 29.3
2:00 272.9 F 29.6
3:00 272.5 F 29.9
4:00 271.5 F 29.3

Min 271.4 F 29.2
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 218.0 F 0.0
10:00 228.1 F 0.0
15:00 222.3 F 0.0
20:00 231.8 F 0.0
25:00 234.2 F 0.0
30:00 233.4 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 21:39
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:17
CYCLE COMPLETE

04/07/2022 02:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____